

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 22 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130303</u>	
Name <u>JEAN ACHEE</u>		ID# <u>1470</u>	Date <u>10-1-2009</u>
A	Agency <u>WASILLA POLICE DEPT.</u>		Phone # <u>352-5401</u>
Instrument Location <u>WASILLA POLICE DEPT.</u>			
B	Alco S/N <u>70108</u>	Target Value <u>.087</u>	High Pressure <u>600</u>
Alco Test Value Average <u>.086</u>		<u>.086</u>	
1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco	
Signature <u>Jean Achee</u>		008 10/6/09	
(OVER)			


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 6<sup>th</sup> day of NOV., 2009.  
Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303

OCTOBER 01, 2009 ✓

OPERATOR'S NAME: JEAN P

ACHEE, JEAN P

OPERATOR'S NUMBER: 1470

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567890

DEPT/AGENCY: WAS1

CASE/REPORT: 09-1001

TEST TYPE: V

ALCO TARGET VALUE: .087 ✓

ALCO S/N: 70108

### --- BREATH ANALYSIS ---

.087 ADJUSTED FOR 29.90 in  
ALCO TARGET .086  
BLANK TEST .000  
INTERNAL STANDARD VERIFIED  
ALCO TV 29.88 in .086 ✓  
BLANK TEST .000 ✓  
SUBJECT SAMPLE .000 ✓  
BLANK TEST .000 ✓  
ALCO TV 29.90 in .086  
BLANK TEST .000

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303

OCTOBER 01, 2009 ✓

TIME 17:38

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.90 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~△